



Certificate of Analysis

1.15333.2500 Water for chromatography (LC-MS Grade) LiChrosolv®
Batch Z0911433

Batch Values

Spec. conductance at 25 °C (at the time of manufacturing)	≤ 1	µS/cm
Al (Aluminium)	≤ 10	ppb
Ca (Calcium)	≤ 100	ppb
Fe (Iron)	≤ 5	ppb
K (Potassium)	≤ 10	ppb
Mg (Magnesium)	≤ 20	ppb
Na (Sodium)	≤ 200	ppb
Any further metal (ICP-MS)	≤ 5	ppb
TOC	≤ 30	ppb
Evaporation residue	≤ 5	mg/l
Fluorescence (as quinine at 254 nm)	≤ 1	ppb
Fluorescence (as quinine at 365 nm)	≤ 0.5	ppb
Gradient grade (at 210 nm)	≤ 5	mAU
Gradient grade (at 254 nm)	≤ 0.5	mAU
Gradient grade (basic absorption (210 nm))	≤ 20	mAU

Batch Values

Anions (IC)		
Chloride (Cl)	≤ 10	ppb
Nitrate (NO ₃)	≤ 10	ppb
Phosphate (PO ₄)	≤ 10	ppb
Sulfate (SO ₄)	≤ 10	ppb

Batch Values

Suitable for LC-MS; Intensity of background mass peak based on reserpine (APCI/ESI positive)	≤ 1	ppb
Suitable for LC-MS; Intensity of background mass peak based on reserpine (APCI/ESI negative)	≤ 20	ppb

Filtered by 0.2µm filter

Date of release (DD.MM.YYYY) 20.02.2024
Minimum shelf life (DD.MM.YYYY) 28.02.2027

Jeannette David
Responsible laboratory manager quality control

This document has been produced electronically and is valid without a signature.